



## Original Research Article

## Comparison of automated method and photometric cyanmethemoglobin method for haemoglobin estimation

Dileep Kumar Jain<sup>1,\*</sup><sup>1</sup>Dept. of Pathology, Hind Institute of Medical Sciences, Safedabad, Barabanki, Uttar Pradesh, India

## ARTICLE INFO

## Article history:

Received 10-08-2020

Accepted 14-08-2020

Available online 03-09-2020

## Keywords:

Hemoglobin

Manual

Automated

Photometer

Cell counter

## ABSTRACT

**Background:** Haemoglobin is one of the commonest tests required in pathology labs. There are many methods of haemoglobin estimation. Common methods used are sahli's acid hematin method, photometric cyanmethemoglobin method with hemoglobin estimation by colorimeter, automated method by cell counter.

**Objective:** To evaluate accuracy of hemoglobin estimation by manual photometric method versus automated method by hematology cell counters using the same sample at the same time.

**Materials and Methods:** Blood samples of 460 adult patients and 72 children (<15 Yrs), including outdoor and indoor, between May 2019 to Aug 2019, attending H.I.M.S. were collected in EDTA tubes. Samples were properly mixed on blood shaker. Hemoglobin estimation done by photometric cyanmethemoglobin method by AIMIL digital colorimeter and by Mindray (BC5150) automated cell counter.

**Results:** Results by photometric method showed higher mean value compared to automated method by 2.52%. Commercial control results showed 2.2% coefficient of variation by Mindray cell counter and 2.9% by photometer.

**Conclusion:** Both methods are accurate with 2.52% more mean value in photometric method. When Haemoglobin is the only test required, photometric method is cost effective and feasible. If sample size is large as in tertiary care hospitals and multiple parameters like complete blood count (CBC) are needed, automated method is time effective and feasible.

© 2020 Published by Innovative Publication. This is an open access article under the CC BY-NC license (<https://creativecommons.org/licenses/by-nc/4.0/>)

### 1. Introduction

Hemoglobin comprises of four globin protein subunits, each having one polypeptide chain and one heme group. Oxygen binds reversibly to the ferrous iron atom in each heme group.<sup>1,2</sup> Main function of haemoglobin is to transport oxygen from lungs to the tissues, where oxygen is utilized for metabolism,<sup>3</sup> mainly to facilitate oxidative phosphorylation in mitochondria.<sup>4</sup> Anaemia is very common in our country, worldwide it affects over 800 million women and children,<sup>5</sup> also there are fair number of cases in advanced countries like US,<sup>6</sup> so haemoglobin is advised frequently to the patients especially in antenatal patients because of increased risk of complications.<sup>7</sup> Hb

measurement is also essential for screening the eligibility of donors in blood banks.<sup>8</sup> The reference range for normal hemoglobin (according to WHO) is 13-18 gm/dl for men, 12-16 gm/dl for women.<sup>9,10</sup>

Hemoglobin being one of the commonest investigations in laboratory, we require economical, feasible and accurate method.<sup>11</sup> Many methods are commonly used, sahli's acid hematin method, cyanmethemoglobin method by photometer, automated method by cell counters.<sup>3</sup> Different lab uses different methods depending on laboratory location, number of patients, availability of technical staff, problem of electricity, affordability. There is variation in hemoglobin measurement done by different instruments due to many reasons including type of sample.<sup>10</sup>

Photometric cyanmethemoglobin method is based on principle that Drabkin reagent reacts with hemoglobin

\* Corresponding author.

E-mail address: [dileepjain64@gmail.com](mailto:dileepjain64@gmail.com) (D. K. Jain).

in blood to form cyanmethemoglobin and developed colour is measured by photometry at 550 nm. Automated method on Mindray cell counter is based on principle of electrical impedance for cell counting and colorimetry for HB estimation.<sup>12</sup> Automated method uses a non cyanide hemoglobin method.<sup>13</sup>

## 2. Aims and objective

The study were to evaluate accuracy, cost effectiveness, suitability and feasibility of photometric versus automated method of haemoglobin estimation.

## 3. Materials and Methods


We took 532 patient's (460 adults and 72 children) Hb estimation into consideration between May 2019 to Aug 2019 including indoor & outdoor patients attending Hind Institute of Medical Sciences, Safedabad, Barabanki, Uttar Pradesh. Adults were between 15 to 85yrs of age. Children were less than 15yrs upto 03 hrs old, newly born baby. 2 ml blood samples were collected in tubes containing K3EDTA anticoagulant. After proper mixing on blood shaker Hb estimation was done simultaneously both by Mindray 5 part hematology cell counter (BC-5150) and digital colorimeter by Aimil. For manual method 5 ml Drabkin's reagent was taken in test tube and 20 ul of blood was mixed and then waited for 5minutes at room temperature before taking absorbance on colorimeter at 550 nm filter against Drabkin's solution. A standard curve is used to know the Hb concentration in the sample by measuring absorbance compared to standard control. Drabkin's solution contains cyanide which is hazardous during handling and disposal.<sup>5</sup> Drabkin's reagent was used from Arkray (formerly Span Diagnostics). Standards were also used from same company. For automated cell counter, control blood was used from Diagon. Standards (control) were run regularly. We did not include hemolyzed sample in our study.

## 4. Results

A total of 532 samples were processed for hemoglobin estimation. Out of these 238 were males and 294 females, 460 adults, 72 children. Mean haemoglobin concentration on Mindray cell counter for adults and children were  $10.62 \pm 2.26$  and  $10.12 \pm 2.41$  respectively. Mean haemoglobin concentration by cyanmethemoglobin method on photometer (AIMIL) for adults and children were  $10.84 \pm 2.05$  and  $10.42 \pm 2.32$  respectively. This shows mean by photometric method is greater by 2.07% for adults and 2.96% for children. Accuracy of both the methods was compared using Microsoft Excel software.

This showed 2.52 % (overall) increase in results by manual method. Comparison of different parameters is tabled.

Variable	Mindray Cell Counter (BC-5150)	Aimil Digital Photometer
Number of patients	532	532
Mean Hb of Adults	10.62	10.84
Mean Hb of Children	10.12	10.42
Lowest Hb among Adults	3.40	4.10
Highest Hb among Adults	16.80	15.10
Lowest Hb among Children	4	4.5
Highest Hb among Children	18.9	19.1
SD (Adults)	2.26	2.05
SD(Children)	2.41	2.32
Costof instrument	Rs 400,000	Rs 9,500
Coefficient of Variation of Controls	2.2%	2.9%
Reagent Stability	Stable	Stable
Technical Skill	Not required	Required



It takes 5 minutes by automated method while 15-20 minutes by photometric method. Photometer requires almost negligible maintenance while automated analyser regular maintenance.

## 5. Discussion

For a developing country like India, economically suitable methodology is the need of the hour. At our institute cost of the test by automated cell counter is Rs 100 whereas cost of manual photometric method is Rs 10 per test. Cost of Mindray Cell Counter is about Rs 4.0 lakh while AIMIL photometer cost only for Rs 9,500. When haemoglobin is the only test required, photometric method is cost effective and feasible. Automated method should be used where complete blood count (CBC), or multiple parameters are required.<sup>14</sup> International committee for standardization in haematology (ICSH) has recommended the 'Drabkin' as the method of choice and have suggested all the other method should be adjusted to be comparable to this method<sup>15</sup> because of availability of internationally accepted reference



**Fig. 1:** Blood Shaker



**Fig. 2:** Mindray Cell Counter BC 5150



**Fig. 3:** Hb Controls of Cell Counter



**Fig. 4:** Photometer (AIMIL)



**Fig. 5:** Hb Standard for photometer

standard calibrator.<sup>16,17</sup> Periodic standards (controls) must be run to maintain accuracy both in manual & automated method. Hemoglobin estimation by automated cell counters is the next best method.<sup>18</sup> There is a significant and positive correlation between the manual and the automated method.<sup>19</sup> For accuracy proper sample collection and proper technique is very important.<sup>20</sup>

## 6. Conclusion

Both methods of Hb estimation are accurate and for single parameter manual method is very cost effective.

## 7. Abbreviations

Hb: Hemoglobin

HIMS: Hind Institute of Medical Sciences

## 8. Source of Funding

None.

## 9. Conflict of Interest

None.

## References

- Marengo-Roweaj. Structure-function relations of human hemoglobins. *Inbaylor Univ Med Center Proc.* 2006;19:239–45.
- Lukin JA, Ho C. The structure– function relationship of hemoglobin in solution at atomic resolution. *Chem Rev.* 2004;104:1219–30.
- Waqar A, Shahida P, Saadat P. Comparison of photometric cyanmethemoglobin and automated methods for hemoglobin estimation. *National Centre Biotechnol Inf.* 2002;14:22–3.
- Thomas C, Lumbab. Physiology of haemoglobin. *Continuing Educ Anaesth Crit Care Pain.* 2012;12:251–6.
- Karakochuk CD, Hess SY, Moorthy D, Namaste S, Parker ME, Rappaport AI, et al. Measurement and interpretation of hemoglobin concentration in clinical and field settings: a narrative review. *Ann N Y Acad Sci.* 2019;1450(1):126–46.
- Le C. The Prevalence of Anemia and Moderate-Severe Anemia in the US Population (NHANES 2003-2012). *PLoS ONE.* 2016;11(11):166635.
- Swaminathan S, Thomas T, Boseb K. Point of care haemoglobin estimation. *Indian J Comm Health.* 2018;30:108–14.
- Whithead RD, Mei Z, Mapango C, D ME. Jefferds. *Ann N Y Acad Sci.* 2019;1450(1):147–51.
- Beutlere W. The denition of anemia: what is the lower limit of normal of the blood hemoglobin concentration. *Blood.* 2006;107(5):1747–50.
- Price CP, Smith I, den Bruel V. Improving the quality of point-of-care testing. *Family Pract.* 2017;35(4):358–64.
- Anchinmanevt S. Evaluation of hemoglobin estimation with non-cyanide alkaline haematin d-575 method. *Int J Res Med Sci.* 2016;4:4297–9.
- Greer JP. Wintrobe, wintrobe’s clinical hematology. vol. 1;.
- Toppo M, Pal DK, Gour D, Melwani V, Dubey M, Mishra A. Comparison of Performance of Digital Hemoglobinometer over Automated Hematology Analyzer for Hemoglobin Estimation and Its user-friendliness among the Pregnant Women in Selected District Hospitals of Madhya Pradesh. *Indian J Community Med.* 2019;44(1):31–4.
- Madhura W, Nitin W, Rakhi T, Agrawal VP. Validation of different tests for haemoglobin estimation. *Int J Biomed Adv Res.* 2014;5(1).
- Balasubramaniam P, Malathi A, et al. Nicotine mediated activation of Pak1/NFkB cascade in pancreatic cancer cells – A pilot study. *Eur J Mol Clin Med.* 1992;38(1):8–9.
- Srivastava T, Negandhi H, Neogi SB, Sharma J, Saxena R. Methods for Hemoglobin Estimation: A Review of “What Works. *J Hematol Transfus.* 2014;2(3):1028.
- Vinaya B, VPuranik G, Bipin SS. Evaluation of non cyanide methods for hemoglobin estimation. *Indian J Pathol Microb.* 2011;54:764–8.
- Shah N, Osea EA, Martinez GJ. Accuracy of noninvasive hemoglobin and invasive point-of-care hemoglobin testing compared with a laboratory analyzer. *Int J Labor Hematol.* 2014;36(1):56–61.
- Babadoko A, Ibrahim I, Musa A, Usman N. Reproducibility of hematological parameters: Manual versus automated method. *Sub-Saharan Afr J Med.* 2016;3(2):65–70.
- Burger S, Pierre-Louis J. A procedure to estimate the accuracy and reliability of hemocue™ measurements of survey workers. Washington: ils. 2003;.

## Author biography

**Dileep Kumar Jain** Associate Professor

**Cite this article:** Jain DK. Comparison of automated method and photometric cyanmethemoglobin method for haemoglobin estimation. *IP J Diagn Pathol Oncol* 2020;5(3):273-276.